

Macro-Distillation

(Note that a condenser is used as the vertical distillation column & no water is attached here; note that the 3-way adaptor at the top of the vertical distillation column has a glass stopper in it – thermometers are not used in our macro distillation set-ups as the temperatures are hugely incorrect; “water in” connection at the bottom of condenser, “water out” connection at the top of the condenser with hose in sink; may need to clamp thin and thick water hoses to prevent leakage; DO NOT hook up vacuum hose from faucet)

